EU Declaration of Conformity
EU Konformitätserklärung
Déclaration UE de conformité

Company **Pyromation LLC, 5211 Industrial Road, Fort Wayne, IN 46825 USA**

declares as manufacturer under sole responsibility, that the product
erklärt als Hersteller in alleiniger Verantwortung, dass das Produkt
déclare sous sa seule responsabilité en qualité de fabricant que le produit

Product **HL06** – Ex db, Ex tb (Flameproof, dust-protected) temperature measurement assemblies with HL06 part number prefix. **Config Code:** XP01, XP02, XP03, XP04, XP05, XP06, XP07, and;
HL08 – Ex ia (Intrinsic Safety) temperature measurement assemblies with HL08 part number prefix.
Config Code: EL01, EL02, EL03, EL04, EL05, EL06, EL07.

Regulations conforms to following European Directives:
den folgenden Europäischen Richtlinien entspricht:
est conforme aux prescription des Directives Européennes suivantes:

ATEX	2014/34/EU	
EMC	2014/30/EU	With transmitter, Without Bluetooth option
RED	2014/53/EU	With transmitter, With Bluetooth option

Standards applied harmonized standards or normative documents:
angewandte harmonisierte Normen oder normative Dokumente:
normes harmonisées ou documents normatifs appliqués:

ATEX	EN IEC 60079-0	(2018)	EN 60079-1	(2014)	Ex db
			EN 60079-31	(2014)	Ex tb
			EN 60079-11	(2012)	Ex Ia
EMC	EN 61010-1	(2010)	EN 61326-2-3	(2018)	
	EN 61326-1	(2013)	EN 61326-2-5	(2018)	
RED	EN 300 328 V2.2.1	(2016)	EN 301 489-1 V2.2.0	(2017)	
	EN 301 489-17 V3.2.0	(2017)	EN IEC 63000	(2018)	

Marking CE 2813 Ex II 2 GD

Ex db IIC T6...T4 Gb, Tamb – see nameplate
Ex tb IIICT+60°C... T110°C Db, Tamb – see nameplate

II 1 G Ex ia IIC T6…T4 Ga, Gb, Ta: - see nameplate
II 2 D Ex ia IIIC T60°C…T100°C Db; Ta: see nameplate

Certification

EU-Type Examination Certificate No. EU-Baumusterprüfbescheinigung Nr. Numéro de l'attestation d'examen UE de type	SIRA 18ATEX1250X	Ex db, Ex tb
	CSANe 21ATEX2080X	Ex ia
Issued by / Ausgestellt von / Délivré par	CSA Group Netherlands B.V. (2813)	
Quality assurance / Qualitätssicherung / Système d'assurance qualité	SIRA 15 ATEX M707 DNV GL (76226) *Applies to EMC, RED*	

Fort Wayne, 05.10.2023
Pyromation LLC

Eric Sawyer
General Manager / Hauptgeschäftsführer / Directeur général