

CALIBRATION LABORATORIES

NVLAP LAB CODE 200502-0


SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

<p>Pyromation Inc. Metrology Laboratory 5211 Industrial Road Fort Wayne, IN 46825-5152 Mr. Chris Moritz Phone: 260-484-2580 Fax: 260-482-6805 E-mail: chris@pyromation.com URL: http://www.pyromation.com</p>	<p>Fields of Calibration Electromagnetics - DC/Low Frequency Thermodynamic</p>
--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
ELECTROMAGNETICS – DC/LOW FREQUENCY			
DC RESISTANCE and CURRENT (20/E05)			
DC Resistance – Generate	0 Ω to 10 Ω	0.005 Ω	
	10 Ω to 100 Ω	0.014 Ω	
	100 Ω to 1 kΩ	0.12 Ω	
	1 kΩ to 10 kΩ	1.2 Ω	
	10 kΩ to 100 kΩ	0.44 kΩ	
	100 kΩ to 1 MΩ	2.3 kΩ	
	1 MΩ to 10 MΩ	23 kΩ	
	10 MΩ to 100 MΩ	0.23 MΩ	
	100 MΩ to 1 GΩ	2.3 MΩ	
DC Resistance – Measure	0 Ω to 10 Ω	0.004 Ω	
	10 Ω to 100 Ω	0.005 Ω	
	100 Ω to 1 kΩ	0.015Ω	
	1 kΩ to 10 kΩ	0.13 Ω	
	10 kΩ to 100 kΩ	1.2 Ω	
	100 kΩ to 1 MΩ	0.37 kΩ	
	1 MΩ to 10 MΩ	0.79 kΩ	
	10 MΩ to 100 MΩ	2.2 kΩ	
	100 MΩ to 1 GΩ	5.8 MΩ	
DC Current – Generate	0 mA to 100 mA	0.007 mA	

2021-09-10 through 2022-09-30
 Effective dates


 For the National Voluntary Laboratory Accreditation Program


CALIBRATION LABORATORIES

NVLAP LAB CODE 200502-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
DC Current – Measure	0 mA to 25 mA 25 mA to 100 mA	0.002 mA 0.005 mA	
DC VOLTAGE (20/E06)			
DC Voltage – Generate	0 mV to 100 mV	0.01 mV	
DC Voltage – Measure	0 mV to 100 mV 100 mV to 1 V 1 V to 10 V	0.002 mV 0.012 mV 0.094 mV	
THERMODYNAMIC			
LABORATORY THERMOMETERS, DIGITAL AND ANALOG (20/T03)			
Thermometer by Comparison	-196 °C -80 °C to 420 °C 420 °C to 660 °C	0.02 °C 0.04 °C 0.49 °C	PRT Type S Reference
	-196 °C -80 °C to 420 °C	0.006 °C 0.021 °C	SPRT
Fixed Point	0.01 °C	0.0034 °C	TPW
Fixed Point	29.7646 °C	0.0033 °C	MPGa
THERMOCOUPLE & PYROMETER INDICATORS (20/T08)			
Digital/Analog Temperature Indicators - Simulated			
Type PT100	-200 °C to 660 °C	0.03 °C	
E	-270 °C to 1000 °C	0.07 °C	
J	-210 °C to 1200 °C	0.06 °C	
K	-270 °C to 1372 °C	0.08 °C	
N	-270 °C to 1300 °C	0.11 °C	
T	-270 °C to 400 °C	0.08 °C	
B	0 °C to 1820 °C	0.14 °C	
R	-50 °C to 1768 °C	0.36 °C	

2021-09-10 through 2022-09-30
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 200502-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
S	-50 °C to 1768 °C	0.36 °C	
Digital/Analog Temperature Indicators - Measure			
Type PT100	-200 °C to 660 °C	0.03 °C	
E	-270 °C to 1000 °C	0.07 °C	
J	-210 °C to 1200 °C	0.06 °C	
K	-270 °C to 1372 °C	0.08 °C	
N	-270 °C to 1300 °C	0.10 °C	
T	-270 °C to 400 °C	0.07 °C	
B	0 °C to 1820 °C	0.11 °C	
R	-50 °C to 1768 °C	0.31 °C	
S	-50 °C to 1768 °C	0.31 °C	
THERMOCOUPLES (20/T11)			
Base Metal by Comparison			
Type E	-196 °C	0.50 °C	PRT
	-75 °C	0.50 °C	
	-40 °C to 420 °C	0.50 °C	
	400 °C to 1000 °C	0.73 °C	Type S Reference
Type J	-196 °C	0.64 °C	PRT
	-75 °C	0.64 °C	
	-40 °C to 420 °C	0.64 °C	
	400 °C to 1100 °C	0.81 °C	Type S Reference
	1100 °C to 1204 °C	1.0 °C	
Type K	-196 °C	0.65 °C	PRT
	-75 °C	0.64 °C	
	-40 °C to 420 °C	0.64 °C	
	400 °C to 1100 °C	0.81 °C	Type S Reference
	1100 °C to 1204 °C	1.0 °C	
	1100 °C to 1372 °C	1.6 °C	Type B Reference
Type N	-196 °C	0.67 °C	PRT
	-75 °C	0.65 °C	
	-40 °C to 420 °C	0.65 °C	

2021-09-10 through 2022-09-30

Effective dates



For the National Voluntary Laboratory Accreditation Program


CALIBRATION LABORATORIES

NVLAP LAB CODE 200502-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC)^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
	400 °C to 1100 °C	0.82 °C	Type S Reference
	1100 °C to 1204 °C	1.0 °C	Type B Reference
	1100 °C to 1300 °C	1.3 °C	
Type T	-196 °C	0.32 °C	PRT
	-75 °C	0.31 °C	
	-40 °C to 420 °C	0.30 °C	
Noble Metal by Comparison Type B	200 °C to 420 °C	0.96 °C	PRT
	400 °C to 1100 °C	0.56 °C	Type S Reference
	1100 °C to 1204 °C	0.81 °C	Type B Reference
	1100 °C to 1450 °C	1.7 °C	
Type R	-40 °C to 420 °C	0.65 °C	PRT
	400 °C to 1100 °C	0.76 °C	Type S Reference
	1100 °C to 1204 °C	0.96 °C	Type B Reference
	1100 °C to 1450 °C	1.7 °C	
Type S	-40 °C to 420 °C	0.57 °C	PRT
	400 °C to 1100 °C	0.76 °C	Type S Reference
	1100 °C to 1204 °C	0.96 °C	Type B Reference
	1100 °C to 1450 °C	1.7 °C	
END			

2021-09-10 through 2022-09-30
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 200502-0

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of $k = 2$. However, laboratories may report a coverage factor different than $k = 2$ to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.5 of NIST Handbook 150, Procedures and General Requirements.


Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Uncertainty values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

2021-09-10 through 2022-09-30

Effective dates



For the National Voluntary Laboratory Accreditation Program